



## VERIFICATION OF COMPLIANCE

*This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.*

**Technical Standard: EMC DIRECTIVE 2014/30/EU  
(EN55022 / EN55024)**

### General Information

Applicant: ONYX Healthcare Inc.  
2F., No.135, Ln. 235, Baoqiao Rd., Xindian Dist.,  
New Taipei City 231, Taiwan (R.O.C.)

### Product Description

EUT Description: Medical Station  
Brand Name: ONYX Healthcare  
Model Number: R-VENUS-222ET-0001; xxVENUS-222xxxxxxxxxx (Where "y" is E & S, "x" is 0~9 ,A~Z,- or blank)

### Measurement Standard

EN 55022: 2010 / AC: 2011

EN 61000-3-2: 2014

EN 61000-3-3: 2013

EN 55024: 2010

(IEC 61000-4-2: 2008; IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010; IEC 61000-4-4: 2012;  
IEC 61000-4-5: 2014; IEC 61000-4-6: 2013; IEC 61000-4-8: 2009; IEC 61000-4-11: 2004)

### Measurement Facilities

Xindian Lab.: Compliance Certification Services Inc.  
No.163-1, Zhongsheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan  
Tel: +886-2-22170894 / Fax: +886-2-22171029

*This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T160426D09-E*

Sam Hu / Assistant Manager

Date: May 6, 2016